

Caution for Unpacking

The two rubber table stoppers are placed under the reagent table. Be sure to remove them before use. If the analyzer is operated without removing them, it may cause the serious failure. Please keep the table stoppers and attach them whenever the analyzer is transported.

